

TEST REPORT: 7191236532-CHM20-01B-RC

Date: 04 MAY 2020

Tel: +65 68851345 Fax: +65 67732912

Client's Ref:

Email: Randy.CHIN@tuv-sud-psb.sg

Note: This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.



PSB Singapore

Add value.
Inspire trust.

SUBJECT

Bacterial Filtration Efficiency (BFE)

CLIENT

JSB TECH PTE LTD
100 Pasir Panjang Road
#06-02 100 Pasir Panjang
118518 Singapore
Singapore

Attn : Ms Paisley Ang

SAMPLE SUBMISSION DATE / TEST DATE

22 Apr 2020 / 29 Apr 2020

DESCRIPTION OF SAMPLE

One sample labelled as "eGee-Pro ® Mask" was submitted by the above company.



METHOD OF TEST

ASTM F2101-19 Standard Test Method for Evaluating the Bacterial Filtration Efficiency (BFE) of Medical Face Mask Materials, Using a Biological Aerosol of Staphylococcus aureus.

Flowrate: 28.3 ± 0.3 L/min

Mean particle size of the challenge aerosol: 3 µm ± 0.3 µm

Test area: Approximately 50 cm²



Laboratory:
TÜV SÜD PSB Pte. Ltd.
No.1 Science Park Drive
Singapore 118221

Phone : +65-6885 1333
Fax : +65-6776 8670
E-mail: enquiries@tuv-sud-psb.sg
www.tuv-sud-psb.sg
Co. Reg : 199002667R

Regional Head Office:
TÜV SÜD Asia Pacific Pte. Ltd.
1 Science Park Drive, #02-01
Singapore 118221
TUV®